

WALKING STICK WOF CHECK

Height: Is the stick the right height for the user. If the stick is too high, the arm will be bent at almost right angles and will not offer suitable support for the user.

Tip: If the tip on the stick is worn, it will not grip the surface properly and you run the risk of slipping and falling.

Handle: Some sticks have foam grips. If this is split or worn, it can be replaced.

Bends and cracks: The stick should be checked to see if it has bent or cracked during prolonged use.

Tightening: Height adjustable walking sticks have a tightening collar. Is this in place and working properly.

Bring your walking stick in to us at Access Mobility and we will check it for you free of charge.

Showrooms:

Access Mobility – Blenheim

Cnr Stuart & Main Sts, Blenheim 7201

Ph: 03 5788002

Access Mobility - Richmond

Cnr McGlashen Avenue & Croucher Street, Richmond, Nelson 7020

Ph: 03 544 7717